



EU DECLARATION OF CONFORMITY (DoC)

Manufacturer:	Becton Dickinson Infusion Therapy Systems Inc., 9450 South State Street, Sandy, Utah 84070, USA
Manufacturer SRN:	US-MF-000017719
Authorised Representative:	Becton Dickinson Ireland Ltd. Donore Road, Drogheda, Co. Louth A92 YW26, Ireland
Authorised Representative SRN:	IE-AR-000007610
Product:	BD Saf-T-Intima™ Safety System
Basic UDI-DI:	038290ESLWZLJRKH
Risk Class and Rule:	Class IIa, Annex VIII, Rule 7
Intended Purpose	<p>BD Saf-T-Intima™ Safety System (PVC with DOA/Citrate) Vascular Access Only (REF 383338, 383339, 383348):</p> <p>The BD Saf-T-Intima™ Safety System is intended to be inserted into a patient's vascular system for short term use to sample blood, monitor blood pressure, or administer fluids intravenously. These catheters may be used for any patient population with consideration given to adequacy of vascular anatomy, appropriateness for solutions being infused and duration of therapy.</p> <p>BD Saf-T-Intima™ Safety System (PVC with DOA/Citrate) Subcutaneous and Vascular Access (REF 383318, 383319, 383328, 383329):</p> <p>The BD Saf-T-Intima™ Safety System is intended to be:</p> <ol style="list-style-type: none">1. Inserted into a patient's vascular system for short term use to sample blood, monitor blood pressure, or administer fluids intravenously.2. Inserted into a patient's subcutaneous tissue for short term use for the administration of fluids and medications. <p>These catheters may be used for any patient population with consideration given to adequacy of vascular anatomy or subcutaneous tissue, appropriateness for solutions being infused and duration of therapy.</p>
Notified Body:	BSI Say Building, John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands Notified Body Number: 2797

We, as the manufacturer of the device(s) take sole responsibility for and hereby declare that the above-mentioned product(s) meet(s) the provisions of the following Directives/ Regulation(s):



- Regulation (EU) 2017/745 of the European Parliament and of the Council on Medical Devices

Conformity Assessment Route:

<input checked="" type="checkbox"/> ANNEX IX Chapter I and III – Quality management System	EC CERTIFICATE No.: MDR 731353
<input type="checkbox"/> ANNEX IX Chapter II - Technical Documentation	EC CERTIFICATE No.:
<input type="checkbox"/> ANNEX X Type Examination	EC CERTIFICATE No.:
<input type="checkbox"/> ANNEX XI Part A Production Quality Assurance	EC CERTIFICATE No.:
<input type="checkbox"/> ANNEX XI Part B Product Verification	EC CERTIFICATE No.:
<input type="checkbox"/> ANNEX II & III Technical Documentation	N/A

Common Specifications (CS):

Number: <Version/Year>	Title:	Full or Partial Application: <Justification>
N/A	N/A	N/A

Devices Covered by this DoC: *<only complete if more than one device is covered by this DoC>*

SKU#	Device Name	Device Class
383318	BD Saf-T-Intima™ Safety System	Class IIa
383319	BD Saf-T-Intima™ Safety System	Class IIa
383328	BD Saf-T-Intima™ Safety System	Class IIa
383329	BD Saf-T-Intima™ Safety System	Class IIa
383338	BD Saf-T-Intima™ Safety System	Class IIa
383339	BD Saf-T-Intima™ Safety System	Class IIa
383348	BD Saf-T-Intima™ Safety System	Class IIa

**Authorised Signatory:**

Name & Title:	Christopher Rogers, VP Regulatory Affairs
On behalf of:	Becton Dickinson Infusion Therapy Systems Inc.
Place of Issue:	9450 South State Street, Sandy, Utah 84070, USA
Date of Issue:	07-Mar-2024
Signature:	<div><div>DocuSigned by: <i>Christopher Rogers</i></div><div><div></div><div>Signer Name: Christopher Rogers Signing Reason: I approve this document Signing Time: 07-Mar-2024 2:36:49 PM PST 36DFBDC7D93A4EDD8A95BFA0996E41F6</div></div></div>

DECLARATION OF CONFORMITY Revision History:

Version:	Detailed Change Description:
A	Original Release
B	Correct conformity assessment route (removing MDR certificate and unselecting Annex IX Chapter II). Remove references to Regulation (EU) 207/2012 on electronic instructions for use of medical devices

**TEMPLATE Revision History:**

Rev	Revision Description	ECO Number	Requested By
06	Removed Certificate Expiration Date from Conformity Assessment Route section of the DoC. This is not required by 2017/745 and does not impact conformity assessment requirements. Modified European Authorized Representative Example in instructions from BD Switzerland to BD Ireland Limited.	500000325481	David Pieratos
05	Updated Authorized Signatory section to include a box with the statement "On behalf of" as well as provide guidance/instructions. This requirement MDR requirement for the DoC was missed in the Revision 4 update.	500000285045	Terri Krutz
04	Updated to include Chapter III in conformity assessment route option "ANNEX IX Chapter I – Quality management System" for all languages. Modified header to include Version Number as some businesses use SAP and others may use other approval and storage systems	500000283041	C. Pell
03	Updated to include Intended Purpose and guidance. Updated Revision History in Footer.	500000230219	David Pieratos
02	Based on recommendations from the BDX European Regulatory Affairs team, the DoC was reformatted to simplify the content to be in line with 2017/745 and MedTech Europe Guidance.	500000213116	Denise Oliveira
01	Original release.	500000190393	Jennifer Jaye