

## EC DECLARATION OF CONFORMITY

### 6.18 Declaration of Conformity

The EC declaration of conformity will be signed before the product is placed on the market and may be referred to in the Quality Document Management system as follows:

Document number	Document Title	Content
VENAGRADO-2	Declaration of conformity	The <b>EC declaration of conformity</b> is the written statement and the single declaration drawn up by the manufacturer to demonstrate the fulfilment of the EU requirements relating to a product bearing the CE marking he has manufactured. The declaration is in respect of all Community acts applicable to the product containing all information required for the identification of Community harmonisation legislation to which the declaration relates.

The following standards and regulations were applied in the documents related to this section:

Directive / Guideline	Title
Council Directive 93/42/EEC	14 June 1993 concerning medical devices
MEDDEV 2.4/1 rev 9	Classification of Medical Devices

## EC DECLARATION OF CONFORMITY

*This declaration of conformity is issued under the sole responsibility of Speciality Fibres and Materials Ltd.*

<b>Legal Manufacturer:</b>	Specialty Fibres and Materials Ltd Galaxy House – 31 Herald Way, Binley Industrial Estate, CV3 2RQ, Coventry, United Kingdom
<b>SRN:</b>	N/A
<b>Intended Use:</b>	Suprasorb Liquacel Ag antimicrobial wound dressings can be used for the management of moderately to heavily exuding wounds.
<b>General Product Name/s:</b>	Suprasorb Liquacel Ag
<b>Variants:</b>	As per Annex I – Product Listing
<b>Applicable Standards:</b>	As Per Annex II
<b>Classification:</b>	Class III – Rule 13
<b>GMDN Code/Term:</b>	47045 – Cavity-wound management dressing, antimicrobial
<b>Basic UDI-DI:</b>	N/A
<b>Conformity Assessment Procedure:</b>	EC Declaration of Conformity in accordance with Annex II of the Medical Device Directive certificate number 1984-MDD-21-752 SFM Annex II.3
<b>Notified Body:</b>	Kiwa Belgelendirme Hizmetleri A.Ş. (CE1984) Tepeören Mevkii Ankara Asfaltı Maret Arkası ITOSB 9. Cadde No: 15 Tuzla, Istanbul Turkey
<b>Authorised Representative:</b>	Advena Ltd Tower Business Centre, 2nd Floor, Tower Street, Swatar, BKR 4013 Malta
<b>EC Certificate Number:</b>	1984-MDD-21-769
<b>Expiry Date</b>	27 May 2024

**Signature:**



**Date and Place:** 05 January 2022, Coventry

**Name:** Nyerngoor Korda Hewitt

**Position:** Director of RAQ

## EC DECLARATION OF CONFORMITY

### ANNEX I – Product Listing

Commercial Name	Product Code	Description
Suprasorb Liquacel Ag	FV000241	120gsm 5 x 5 cm
Suprasorb Liquacel Ag	FV000242	120gsm 10 x 10 cm
Suprasorb Liquacel Ag	FV000243	120gsm 10 x 12.5 cm
Suprasorb Liquacel Ag	FV000244	120gsm 15 x 15 cm
Suprasorb Liquacel Ag	FV000245	120gsm 2 x 45 cm
Suprasorb Liquacel Ag	FV000246	160gsm 5 x 5 cm
Suprasorb Liquacel Ag	FV000247	160gsm 10 x 10 cm
Suprasorb Liquacel Ag	FV000248	160gsm 10 x 12.5 cm
Suprasorb Liquacel Ag	FV000249	160gsm 15 x 15 cm
Suprasorb Liquacel Ag	FV000250	160gsm 2 x 45 cm
Suprasorb Liquacel Ag	FV000251	200gsm 5 x 5 cm
Suprasorb Liquacel Ag	FV000252	200gsm 10 x 10 cm
Suprasorb Liquacel Ag	FV000253	200gsm 10 x 12.5 cm
Suprasorb Liquacel Ag	FV000254	200gsm 15 x 15 cm
Suprasorb Liquacel Ag	FV000255	200gsm 2 x 45 cm

### ANNEX II – List of Standards

This present declaration is also in conformity with the following European standards and Common Specifications:

Number	Standards Title	Version
EN ISO 11137-1	Sterilization of health care products - Radiation - Part 1: Requirements for development, validation and routine control of a sterilization process for medical devices	2006/A1: 2013 / 2015
EN ISO 11137-2	Sterilization of health care products - Radiation - Part 2: Establishing the sterilization dose	2013 / 2015
EN ISO 11137-3	Sterilization of health care products - Radiation - Part 3: Guidance on dosimetric aspects	2017

## EC DECLARATION OF CONFORMITY

EN ISO 10993-1	Biological evaluation of medical devices - Part 1: Evaluation and testing within a risk management process	2009/ AC:2010
EN ISO 10993-3	Biological evaluation of medical devices - Part 3: Tests for genotoxicity, carcinogenicity and reproductive toxicity	2014
EN ISO 10993-5	Biological evaluation of medical devices - Part 5: Tests for in vitro cytotoxicity	2009
EN ISO 10993-6	Biological evaluation of medical devices - Part 6: Tests for local effects after implantation	2016
EN ISO 10993-10	Biological evaluation of medical devices - Part 10: Tests for irritation and skin sensitization	2013
EN ISO 10993-11	Biological evaluation of medical devices - Part 11: Tests for systemic toxicity	2009
EN ISO 10993-12	Biological evaluation of medical devices - Part 12: Sample preparation and reference materials	2012
EN ISO 10993-17	Biological evaluation of medical devices - Part 17: Establishment of allowable limits for leachable substances	2002
EN ISO 10993-18	Biological evaluation of medical devices - Part 18: Chemical characterization of materials	2005 / 2009
EN ISO 11607-1	Packaging for terminally sterilized medical devices - Part 1: Requirements for materials, sterile barrier systems and packaging systems	2009 / 2019
EN ISO 11607-2	Packaging for terminally sterilized medical devices - Part 2: Validation requirements for forming, sealing and assembly processes	2006 / 2019
EN ISO 11737-1	Sterilization of medical devices - Microbiological methods - Part 1: Determination of a population of microorganisms on products	2006 / 2018
EN ISO 11737-2	Sterilization of medical devices - Microbiological methods	2009 / 2020

## EC DECLARATION OF CONFORMITY

	- Part 2: Tests of sterility performed in the definition, validation and maintenance of a sterilization process	
EN ISO 13485	Medical devices - Quality management systems - Requirements for regulatory purposes	2016
EN ISO 14155	Clinical investigation of medical devices for human subjects - Good clinical practice	2011
EN ISO 13726-1	Test methods for primary wound dressings - Part 1: Aspects of absorbency	2002 + AC:2003
EN ISO 14971	Medical devices - Application of risk management to medical devices	2019
EN ISO 15223-1	Medical devices - Symbols to be used with medical device labels, labelling and information to be supplied - Part 1: General requirements	2016
ASTM D 4169-16	Standard Practice for Performance Testing of Shipping Containers and Systems	2016
ASTM F1980-16	Standard Guide for Accelerated Aging of Sterile Barrier Systems for Medical Devices	2016
EN 556-1	Sterilization of medical devices. Requirements for medical devices to be designated "STERILE" Requirements for terminally sterilized medical devices	2001
EN 1041	Information supplied by the manufacturer of medical devices	2008
EN 62366-1	Medical devices — Part 1: Application of usability engineering to medical devices	2015
EN 62366-2	Medical devices — Part 2: Guidance on the application of usability engineering to medical devices	2016
EN ISO 14644-1	Cleanrooms and associated controlled environments. Classification of air cleanliness	2015

## EC DECLARATION OF CONFORMITY

EN ISO 14644-2	Cleanrooms and associated controlled environments. Specifications for testing and monitoring to prove continued compliance with ISO 14644-1	2015
----------------	---	------